

DEVELOPMENT OF SMART MULTI-SENSORY SOLUTIONS FOR THE ACTIVE AGEING AND WELL BEING

Student: Ruben Crispino

Tutor: Prof. Bruno Ando'

A.Y.: 2018/2019

Ph.D. Brainstorming Day



**UNIVERSITÀ
degli STUDI
di CATANIA**




Research Topics

Development of smart multi-sensor solution in the framework of assistive technology

- Fall detection and classification
- Postural instability (primary cause for falls)
- Assisted diagnosis system in neurological context
- Navigation aids



TOP 5 CAUSES OF FALLS




Medication—Many drugs (i.e., sedatives, anti-depressants) reduce mental alertness, affect balance and gait and cause drops in systolic blood pressure while standing. Mixing certain medications increases these effects, causing falls. **Solution:** Have a home care professional carefully monitor medications and interactions.

Impaired Vision—Cataracts and glaucoma alter depth perception, visual acuity, peripheral vision and susceptibility to glare. **Solution:** Add color and contrast to identify objects, such as grab bars and handrails.

Weakness, Low Balance—Weakness and lack of mobility leads to many falls. **Solution:** Exercise regularly to boost strength and muscle tone.


Home Hazards—Most homes are full of falling hazards. **Solution:** Add grab bars in the bathroom, install proper railings on both sides of stairways, improve the lighting, remove loose rugs and fix uneven or cracked sidewalks.

Chronic Conditions—Parkinson's, heart disease and other conditions increase the risk of falling. **Solution:** Enlist specially-trained caregivers to ensure that patients follow their treatment plans, assist them to doctor appointments and recognize red flags.




- 30%** 100% of people who fall suffer moderate to severe injuries. These injuries can make it hard to get around or live independently.
- 20%**
- 1 IN 3** adults age 65 and older fall each year, yet more than half talk to their healthcare providers.
- 15** Every 15 seconds across America, a senior citizen is sent to the E.R. for a fall-related injury.

Fall Detection










General Methodology

2



Fall Detection

Signature Generation



Dataset:

- Fs= 500 Hz
- Acquisition duration: 10 s
- Repetitions: 10
- Users: 5

Moving average filter:

- Time window: 100 samples (200 ms)
- Shift: 1 sample (2 ms)


Normalization:

- Every signal is normalized as respect the absolute maximum and minimum

Fall Detection

Correlation Phase

Features: correlation's maximum value between the unknown event and the signature



10 Forward Falls (FF)



	FF	BF	LF	SI
0.9447	0.7318	0.7195	0.7043	
0.8225	0.942	0.7585	0.7585	
0.4387	0.9296	0.5806	0.6925	
0.7013	0.9563	0.7083	0.7494	
0.9564	0.7188	0.7336	0.672	
0.9028	0.3757	0.58	0.4524	
0.9431	0.4865	0.6759	0.4685	
0.9346	0.5353	0.6278	0.6146	
0.9422	0.6452	0.7149	0.5741	
0.8465	0.9086	0.7487	0.7468	

High Cross-correlation

Low Cross-correlation


Very high Cross-correlation

Fall Detection



FP = False Positives
 FN = False Negatives
 TP = True Positives
 TN = True Negatives

Roc Curves



$$Specificity = \frac{TN}{TN + FP}$$

It evaluates the algorithm performance on **evaluating, in a specific way, the type of event.**

$$Sensitivity = \frac{TP}{TP + FN}$$

It evaluated the algorithm performance on **detecting an event.**

Fall Detection

Classification Results

10 Forward Falls (FF)

FF	BF	LF	SI
0.9447	0.7318	0.7195	0.7043
0.8225	0.942	0.7585	0.7585
0.4387	0.9296	0.5806	0.6925
0.7013	0.9563	0.7083	0.7494
0.9564	0.7188	0.7336	0.672
0.9028	0.3757	0.58	0.4524
0.9431	0.4865	0.6759	0.4685
0.9346	0.5353	0.6278	0.6146
0.9422	0.6452	0.7149	0.5741
0.8465	0.9086	0.7487	0.7468

Threshold values computed by means of the ROC curves theory

0.7664 0.7555 0.8155 0.8282

FF	BF	LF	SI
1	0	0	0
1	1	0	0
0	1	0	0
0	1	0	0
1	0	0	0
1	0	0	0
1	0	0	0
1	0	0	0
1	0	0	0
1	0	0	0
1	1	0	0

10 Classified Forward Falls

- Mis-classification
- Multiple Classification

Threshold-Based Algorithm


Index	BF	FF	LF	SI	Av
Sensitivity	50,02	96,66	91,60	100	84,57
Specificity	99,76	96,76	100	75,77	93,07

Advanced Threshold Algorithm

Index	BF	FF	LF	SI	Av
Sensitivity	87,42	94,83	100	100	95,56
Specificity	99,88	95,88	99,26	99,30	98,58

Solution:




Data fusion Analysis on user position after fall



Indoor Navigation

Methodology and Devices

7




Advantages of the proposed solution:

- Continuous US distance measurement
- Autonomous Embedded Architecture
- Real-Time Localization by optimal operation Timing

- Smart Paradigms to improve the system Accuracy
- Multiple contest of applications

Indoor Navigation




$$\begin{bmatrix} u_x \\ u_y \end{bmatrix} = \begin{bmatrix} (A_j)^{-1} & B_j \end{bmatrix} \begin{bmatrix} (d_{2j}^2 - d_{1j}^2) + (x_{1j}^2 - x_{2j}^2) + (y_{1j}^2 - y_{2j}^2) \\ (d_{3j}^2 - d_{1j}^2) + (x_{1j}^2 - x_{3j}^2) + (y_{1j}^2 - y_{3j}^2) \end{bmatrix}$$

with $\begin{cases} j = 1, 2, \dots, M \text{ Number of triad} \\ (x_{ij}, y_{ij}) \text{ The } i\text{-th corner of the } j\text{-th triad} \\ d_{ij} \text{ The user distance from the } i\text{-th corner of the } j\text{-th triad} \end{cases}$

Multi Trilateration

- (x_i, y_i) Environmental Nodes Coordinates with $i = 1, 2, \dots, N$
- (u_x, u_y) User Position
- d_i Distance User / i -th Environmental Node with $i = 1, 2, \dots, N$
- Redundancy of the User Position estimates. The number of combinations is a function of N (Environmental Nodes):


$$M = C(N, 3) = \frac{N!}{(N-3)! 3!}$$




Indoor Navigation

Compensation Models

- **Trasdution Diagram (NODE A)**





- **Calibration Diagram (NODE A)**



Indoor Navigation

Experimental Setup

- **Task of the Characterization:**
 - Accuracy
 - Precision
- **Nominal Positions:**
 - MEASUREMENT GRID
- **Samples collected for each Distance:**
 - Dataset of 100 samples
- **Influence parameter monitored:**
 - Temperature (°C)



Postural instability

11


Objective and Device

General objective:

- To have a system able to:
 - Monitor the user pathology evolution (postural instability is an early symptom of the Parkinson)
 - Validate the pharmacological therapies and rehabilitation strategy.





Implementation:

- A low-cost wearable device and advanced signal processing aimed at the continuous monitoring of the user's postural sway.











Postural instability

Stabilogram



Postural instability







Features



Postural instability




Features Fusion




RFID-based systems

Realized Devices



Orientation system for the visually impaired



User habits monitoring

RFID-based systems

Orientation system: results


Aim:

To validate the assessment results and to quantify the effective utility of the device when used by real users.


Users:

- 6 user in normal physical condition (10 trials/user) **(no end-users)**

With and without the use of the assistive system.




Each value is the sum of 10 trials





Steps/s for 10 trials

RFID-based systems




To evaluate the best power to reading range ratio for the addressed application

User habits monitoring system: results



Assessment of the device behaviour during real use

Projects & Schools & Conferences



PH.D. SCHOOL "ITALO GORINI" 2018
GENEVA, SWITZERLAND, SEPTEMBER 10 - 14, 2018



foritaal
Ambient Assisted Living
9° Forum Italiano – 2-4 Luglio 2018, Lecce

foritaal
Ambient Assisted Living
10° Forum Ancona 2019

FORUM NAZIONALE DELLE MISURE
PERUGIA 2019

MISURE 2018

2019 IEEE
SENSORS APPLICATIONS SYMPOSIUM
MARCH 11-13, 2019 | SOPHIA ANTIPOLIS, FRANCE

Journal

- **Advanced Solutions Aimed at the Monitoring of Falls and Human Activities for the Elderly Population**

B. Andò, S. Baglio, S. Castorina, R. Crispino, V. Marletta, MDPI-Technologies, vol. 7, no. 3, pp. 59

- **A smart inertial system for fall detection (Accepted with minor revision)**

B. Andò, S. Baglio, S. Castorina, R. Crispino, V. Marletta, JAIHC (Journal of Ambient Intelligence and Humanized Computing)

- **A Fluxgate-Based Approach for Ion Beam Current Measurement in ECRIS Beamline: Design and Preliminary Investigations**

B. Andò, S. Baglio, R. Crispino, S. Graziani, V. Marletta, A. Mazzaglia, V. Sinatra, D. Mascali, G. Torrisi, IEEE Transactions on Instrumentation and Measurement, vol. 68, no. 5, pp. 14771484

- **A measurement strategy to assess the optimal design of an RFID-based navigation aid**

B. Andò, S. Baglio, V. Marletta, R. Crispino, A. Pistorio, IEEE Transactions on Instrumentation and Measurement, vol. 68, no. 7, pp. 23562362

Two more under review: MDPI-Sensors, IEEE Transactions on Instrumentation and Measurement

Proceedings

- **NATIFLife-A Smart Sensor Network for Assistive Domotics**

B. Andò, S. Baglio, S. Castorina, R. Crispino, S. Graziani, V. Marletta, L. Porcaro, S. Salupo, 2019 IEEE International Symposium on Measurements & Networking (M&N)

- **A measurement procedure for the optimization of a distributed indoor localization system**

B. Andò, S. Baglio, S. Castorina, R. Crispino, V. Marletta, 2019 IEEE International Symposium on Measurements & Networking (M&N)

- **A sensing platform for postural behavior assessment**

B. Andò, S. Baglio, R. Crispino, V. Marletta, V. Dibilio, G. Mostile, A. Nicoletti, M. Zappia, 2019 IEEE International Symposium on Measurements & Networking (M&N)

- **A measurement strategy to assess performances of Fall Detector paradigms**

B. Andò, S. Baglio, R. Crispino, V. Marletta, 2019 IEEE International Instrumentation and Measurement Technology Conference (I2MTC)

- **An RF-ID based solution of Assistive Technology for User Activity Monitoring**

B. Andò, S. Baglio, S. Castorina, R. Crispino, V. Marletta, 2019 IEEE Sensors Applications Symposium (SAS)

- **An inkjet printed pressure sensor for applications in Active Ageing monitoring**

B. Andò, S. Baglio, S. Castorina, R. Crispino, V. Marletta, 2019 IEEE Sensors Applications Symposium (SAS)

- **A Smart Inertial Pattern for the SUMMIT IoT Multi-platform**

B. Andò, S. Baglio, R. Crispino, L. L'Episcopo, V. Marletta, M. Branciforte, M. C. Virzi, Italian Forum of Ambient Assisted Living, pp. 311-319

- **An Embedded Localization System for the SUMMIT IoT Multi-platform**

R. Crispino, B. Andò, S. Baglio, V. Marletta, Italian Forum of Ambient Assisted Living, pp. 377-384

- **A fluxgate based approach for ion beam current measurement in electron cyclotron resonance ion sources beamline**

B. Andò, S. Baglio, R. Crispino, S. Graziani, V. Marletta, A. Mazzaglia, V. Sinatra, D. Mascali, G. Torrasi, IEEE International Instrumentation and Measurement Technology Conference (I2MTC)

- **A pressure measurement for pull test validation in postural instability assessment**

B. Andò, S. Baglio, R. Crispino, V. Marletta, V. Dibilio, G. Mostile, A. Nicoletti, M. Zappia, 2018 IEEE Sensors Applications Symposium (SAS)

- **Low Cost Inkjet Printed Sensors: From Physical to Chemical Sensors**

B. Andò, S.e Baglio, V. Marletta, R. Crispino, S. Castorina, A. Pistorio, G. Di Pasquale, A. Pollicino, Convegno Nazionale Sensori, pp. 297-308

- **A NeuroFuzzy approach for Fall detection**

B. Ando, S. Baglio, V. Marletta, R. Crispino, International Conference on Engineering Technology and Innovation (ICE/ITMC), pp. 1312-1316

- **Video tutorial for the IEEE Instrumentation and Measurement Society**

Ruben Crispino <http://ieee-ims.org/evts/video-tutorial/strain-gauge-theory-and-application>

